

Certificate of Analysis

Low Gold Material –AuBlank68

Product contains less than 0.004 ppm gold.

Date of certification: 11 November 2013

The above value applies to product in sachets that have the identification number 279 280.

A mixture of finely pulverized feldspars and basalt that had previously been analysed and found to have a low gold content was placed into sachets in a gold-free environment. To the best of our knowledge the component materials have been obtained from areas where no gold has previously been mined and have been crushed and pulverized through machinery that has not been used for the treatment of gold ore. A total of 3822 sachets were filled with 52g of material. 24 sachets were selected at regular intervals from the packaging run and were submitted to a laboratory that specializes in low-level analysis for gold. The samples were analysed by aqua regia digestion followed by enhanced graphite furnace atomic absorption spectroscopy. The average of the 24 analyses was 0.3 parts per billion of gold, with the highest value being 2.0 parts per billion gold.

We certify the material at a level that is at least double the highest gold value obtained in the samples analysed. It is thus certified as less than 0.004 ppm gold.

A copy of the analytical report from the independent laboratory is available on request.

Prepared and Certified By:

Brett Coombridge
Rocklabs Reference Materials
Division of Scott Technology NZ Ltd
12/21 Poland Road, Wairau Valley
Auckland
NEW ZEALAND

Email: b.coombridge@scott.co.nz
Telephone: +64 9 444 3534